



Edge™

Electrodes

REF E1450G

REF E1450X

REF E1450-4

REF E1450-6

REF E1475X

en Instructions for Use

fr Mode d'emploi

pt Instruções de uso

da Brugervejledning

de Gebrauchsanleitung

es Instrucciones de uso

fi Käyttöohjeet

it Istruzioni per l'uso

nl Gebruiksaanwijzing

no Bruksanvisning

sv Bruksanvisning

ru Инструкция по применению

zh 使用说明

en

Edge™

[REF] E1450G	Blade Electrode 3" (7.62 cm)
[REF] E1450X	Hex-locking Blade Electrode 2.5" (6.35 cm)
[REF] E1450-4	Blade Electrode 4" (10.16 cm)
[REF] E1450-6	Blade Electrode 6.5" (16.51 cm)
[REF] E1475X	Hex-locking Blade Electrode 2.75" (7.0 cm)

For use with a max peak voltage of 5600 V.



Not made with natural rubber latex



Do not use if package is opened or damaged

Indications for Use

The Edge Coated Electrodes are intended as an alternative to uncoated stainless steel electrodes for use in conventional monopolar electro-surgical accessories. The coated electrodes are intended for use in situations where monopolar electro-surgical cutting and coagulation are normally used.

Warning

This product cannot be adequately cleaned and/or sterilized by the user in order to facilitate safe reuse, and is therefore intended for single use. Attempts to clean or sterilize these devices may result in a bio-incompatibility, infection, or product failure risks to the patient.

Explosion Hazard Do not use electro-surgery in the presence of flammable anesthetics.

Warning

Fire Hazard Both oxygen (O₂) and nitrous oxide (N₂O) support combustion. Watch for enriched O₂ and N₂O atmospheres near the surgical site, especially during head and neck surgery. Enriched O₂ atmospheres may result in fires and burns to patients or surgical personnel.

Fire/Explosion Hazard The following substances contribute to increased fire and explosion hazards in the operating room:

- Oxygen-enriched environments
- Oxidizing agents, such as N₂O atmospheres
 - Verify all anesthesia circuit connections are leak free before and during use of electrosurgery.
 - Verify endotracheal tubes are leak free and that the cuff seals properly to prevent oxygen leaks.
 - If an uncuffed tube is in use, pack the throat with wet sponges around the uncuffed tube.
 - If possible, stop supplemental oxygen at least one minute before and during use of electrosurgery.

Warning

- Alcohol-based skin prepping agents and tinctures
 - Activate the electrosurgical unit only after vapors from skin prep solutions and tinctures have dissipated.
- Naturally occurring flammable gases (such as methane) that may accumulate in body cavities

Fire Hazard The sparking and heating associated with electrosurgery can provide an ignition source.

Observe fire precautions at all times:

- When using electrosurgery in the same room with gases or flammable substances, prevent pooling of fluids and the accumulation of gases under surgical drapes or near the surgical site.
- Tissue buildup (eschar) on the tip of an active electrode poses a fire hazard, especially in oxygen-enriched environments, such as in throat or mouth procedures. Eschar plus high oxygen may create embers. Keep the electrode clean and free of all debris.

Warning

- Facial and other body hair is flammable. Water-soluble surgical lubricating jelly may be used to cover hair close to the surgical site to decrease flammability.

The electrode must fit completely and securely into the pencil. An incorrectly seated electrode may result in burns to the patient or surgical personnel.

Fire Hazard Always place the active electrode in a clean, dry, insulated safety holster when not in use.

- Electrosurgical accessories that are activated or hot from use can cause unintended burns to the patient or surgical personnel.
- Electrosurgical accessories may cause fire or burns if placed close to or in contact with flammable materials, such as gauze or surgical drapes. Place longer electrodes (such as extended electrodes) away from the patient and drapes.

Warning

Confirm proper electrosurgical settings prior to and during a procedure. Use the lowest power settings to achieve the desired effect. If increased power settings are requested, check the patient return electrode and all accessory connections before major power setting adjustments.

Some surgeons may elect to “buzz the hemostat” during surgical procedures. It is not recommended, and the hazards of such a practice probably cannot be eliminated. Burns to the surgeon’s hands may result. To minimize the risk, take these precautions:

- Do not lean on the patient, the table, or the retractors while buzzing the hemostat.
- Activate cut rather than coag. Cut has a lower voltage than coag.
- Use the lowest power setting possible for the minimum time necessary to achieve hemostasis.
- Activate the generator after the accessory makes contact with the hemostat. Do not arc to the hemostat.

Warning

- Firmly grasp as much of the hemostat as possible before activating the generator. This disperses the current over a larger area and minimizes the current concentration at the fingertips.
- “Buzz the hemostat” below hand level (as close as possible to the patient) to reduce the opportunity for current to follow alternate paths through the surgeon’s hands.
- When using a coated or nonstick blade electrode, place the **edge** of the electrode against the hemostat or other metal instrument.

Precaution

Always use the lowest power setting that achieves the desired surgical effect. Use the active electrode for the minimum time necessary in order to reduce the possibility of unintended burn injury.

Precaution

Pediatric applications and/or procedures performed on small anatomic structures may require reduced power settings. The higher the current flow and the longer the current is applied, the greater the possibility of unintended thermal damage to tissue, especially during use on small appendages.

Before use, examine the electrosurgical unit and accessories for defects. Do not use cables or accessories with damaged (cracked, burned, or taped) insulation or connectors.

Do not modify or add to the insulation of active electrodes.

Activate the electrosurgical unit **only** when ready to deliver electrosurgical current and the active tip is in view (especially if looking through an endoscope).

Deactivate the electrosurgical unit **before** the tip leaves the surgical site.

Precaution

The electrodes are intended for single use only. Safely **discard after use** to prevent injury to hospital personnel. These electrodes are not designed to withstand resterilization. **Do not resterilize.**

Notice

This electrode has a coating to reduce sticking of eschar. Cleaning the electrode with a scratch pad or other abrasive object, scraping with a sharp object, or bending beyond 90 degrees may damage the electrode. If the electrode is damaged, discard it.

Using coated electrodes at high power settings may cause damage to the coating. If the coating is damaged, discard the electrode.


Electrosurgical generators (e.g., Force FX™ or Force EZ™ generators) produce desired surgical effects at lower power cut mode settings than conventional electrosurgical generators. The electrode coating may deteriorate when used with tissue response generators at higher power settings.

Important


Wipe the electrode often with moist gauze or other material.

Instructions for Use

1. Ensure the pencil is not connected to the generator.




2. Grasp the insulating sleeve on the electrode, and insert the electrode into the pencil.



3. Ensure the insulating sleeve fits securely inside the nose of the pencil so that the nose overlaps the insulating sleeve by at least 1/8" (0.3 cm).

4. Hex electrodes have a depth indicator. The line on the depth indicator should be flush with the tip of the handswitching pencil.



5. A tip protector covers the coated end of some electrodes. If a tip protector is present, remove it before use.



STERILE R



Single use

**RX
ONLY**



Do not resterilize



Consult
instructions
for use





CE
0086

Part No. PT00063917

COVIDIEN, COVIDIEN with logo, and Covidien logo and Positive Results for Life are U.S. and internationally registered trademarks of Covidien AG. Other brands are trademarks of a Covidien company, [™] brands are trademarks of their respective owner.

© 2012 Covidien.

 Covidien llc,
15 Hampshire Street, Mansfield, MA 02048 USA.
 Covidien Ireland Limited,
IDA Business & Technology Park, Tullamore.

Made in Mexico. Printed in Mexico.

www.covidien.com

REV 02/2017