

	INSTRUCTIONS FOR USE	
	CURETTES	TF01-IFU-005
	REVISION 1	

Intended use

Curettes are surgical instruments used to remove excess tissues or bony growths, using a scraping or scooping motion. They are available in various sizes and shapes.

Reuse of Single Use Product

This Product is designated as 'Single Use' ONLY. It should be used for one-patient, during a single procedure and then discarded. It is not intended to be reprocessed and used again, even on the same patient. Along with an increased risk of infection, the safety of the product cannot be guaranteed with multiple use.

Product inspection before use

Inspect device for rust, dents, cracks, pits and pores before use.

Warning

- Do not use if product or sterile barrier system is damaged.
- Products supplied non-sterile must be sterilised before use.
- Please follow our sterilisation instructions prior to use.
- Do not use beyond the expiry date stated on the packaging.

Sterilisation

Instruments can be sterilised using the following methods:

Ethylene Oxide

Concentration: up to 450mg/l

Temperature: 60°C

Exposure: up to 6 hours

Humidity: up to 100%

Irradiation

Dose to be validated by user, up to 70kGy.

Contraindications

Instruments should not be used for anything other than their intended use. They are not to be used on patients who have an allergy or sensitivity to stainless steel, or any other component material e.g. Zinc.

Intended User

Our instruments are designed for use by appropriately qualified clinicians, who will have good knowledge of their features and how they should be used. These products are not intended for use by lay persons.

Disposal:

Used instruments should be disposed of in accordance with Local Clinical Waste Guidance.

		 Rociale Healthcare Ltd, Ty Mynydd, Cwm Cynon Business Park (North), Mountain Ash, Rhondda Cynon Taff, CF45 4ER, UK	<table border="1"> <tr> <td>EC</td> <td>REP</td> </tr> </table> Advena Limited, Tower Business Centre, 2nd Flr, Tower Street, Swatar, BKR 4013, Malta	EC	REP
EC	REP				